

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Salmonella PLUS

LOT Lot: 22105

Expiry: 2027-03

REF Art. No.: F5111

specification limit

Control	Cp-value	QC range
Positive Control	30,4	27,1 - 33,7

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	22105	2027-03	o.k.
1	Reaction Mix	22105	2027-03	o.k.
2	Taq Polymerase	22105	2027-03	o.k.
3	Positive Control	22105	2027-03	o.k.

Date 13.03.2025

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.