

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® *Listeria monocytogenes* PLUS

LOT Lot: 21354

Expiry: 2026-08

REF Art. No.: F5113

specification limit

Control	Cp-value	QC range
Positive Control	29,2	25,9 - 32,5

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	21354	2026-08	o.k.
1	Reaction Mix	21354	2026-08	o.k.
2	Taq Polymerase	21354	2026-08	o.k.
3	Positive Control	21354	2026-08	o.k.

Date 27.08.2024

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.