

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Listeria Screening PLUS

LOT Lot: 23115

Expiry: 2027-03

REF Art. No.: F5117

specification limit

Control	Cp-value	QC range
Positive Control	29,6	26,3 - 32,9

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	23115	2027-03	o.k.
1	Reaction Mix	23115	2027-03	o.k.
2	Taq Polymerase	23115	2027-03	o.k.
3	Positive Control	23115	2027-03	o.k.

Date 12.03.2025

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.