

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® EHEC/EPEC 4plex

LOT Lot: 21135

Expiry: 2027-03

REF Art. No.: F5128

specification limit

Control	Cp-value	QC range
Positive Control	26,1	22,8 - 29,4

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	21135	2027-03	o.k.
1	Reaction Mix	21135	2027-03	o.k.
2	Taq Polymerase	21135	2027-03	o.k.
3	Positive Control	21135	2027-03	o.k.

Date 26.03.2025

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.