

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Foodborne Pathogens 4plex

LOT Lot: 22494

Expiry: 2026-12

REF Art. No.: F5175

specification limit

Control	Cp-value	QC range
Positive Control	28,3	23,8 - 32,8

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	22494	2026-12	o.k.
1	Reaction Mix	22494	2026-12	o.k.
2	Taq Polymerase	22494	2026-12	o.k.
3	Positive Control	22494	2026-12	o.k.

Date 05.02.2025

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.