

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® STEC 4plex ONE

LOT Lot: 23095

Expiry: 2027-02

REF Art. No.: F5265

specification limit

| Control | Cp-value | QC range |
|------------------|----------|-------------|
| Positive Control | 29,3 | 26,0 - 32,6 |

Certification

| kit code | description | LOT | Expiry | result certification |
|----------|------------------|-------|---------|----------------------|
| 1 | Reaction Mix | 23095 | 2027-02 | o.k. |
| 1 | Reaction Mix | 23095 | 2027-02 | o.k. |
| 2 | Taq Polymerase | 23095 | 2027-02 | o.k. |
| 3 | Positive Control | 23095 | 2027-02 | o.k. |
| L | Lysis Buffer | 23095 | 2027-02 | o.k. |
| L | Lysis Buffer | 23095 | 2027-02 | o.k. |

Date 10.03.2025

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.