

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Enterobacteriaceae Screening PLUS

LOT Lot: 24044

Expiry: 2026-01

REF Art. No.: F5507

specification limit

Control	Cp-value	QC range
Positive Control	31,0	27,7 - 34,3

Certification

kit code	description	LOT	Expiry	result certification
	Manual	-	-	o.k.
1	Reaction Mix	24044	2026-01	o.k.
1	Reaction Mix	24044	2026-01	o.k.
2	Taq Polymerase	24044	2026-01	o.k.
3	Positive Control	24044	2026-01	o.k.

Date 01.02.2024

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.