

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Norovirus/Hepatitis A 3plex

LOT Lot: 21015

Expiry: 2026-12

REF Art. No.: F7124

specification limit

Control	Cp-value	QC range
Positive Control	27,3	24,0 - 30,6

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	21015	2026-12	o.k.
1	Reaction Mix	21015	2026-12	o.k.
2	PP-Mix	21015	2026-12	o.k.
3	Enzyme Mix	21015	2026-12	o.k.
N	PCR Water	21015	2026-12	o.k.
P	Positive Control	21015	2026-12	o.k.
R	Internal Control RNA	21015	2026-12	o.k.
R	Internal Control RNA	21015	2026-12	o.k.

Date 17.01.2025

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.