

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Hepatitis E PLUS

LOT Lot: 22125

Expiry: 2027-03

REF Art. No.: F7142

specification limit

Control	Cp-value	QC range
Positive Control	27,1	23,8 - 30,4

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	22125	2027-03	o.k.
1	Reaction Mix	22125	2027-03	o.k.
2	PP-Mix	22125	2027-03	o.k.
3	Enzyme Mix	22125	2027-03	o.k.
N	PCR Water	22125	2027-03	o.k.
P	Positive Control	22125	2027-03	o.k.
R	Internal Control RNA	22125	2027-03	o.k.
R	Internal Control RNA	22125	2027-03	o.k.

Date 21.03.2025

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.